

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130132</u>	
A	Name <u>Matthew E. Armstrong</u>	ID# <u>5904</u> <u>MEAU</u> <u>2/10/09</u> <u>CR</u>	Date <u>1 Feb 09</u> ✓
	Agency <u>Seward PD</u>	Phone # <u>(907) 224-3338</u>	
	Instrument Location <u>410 Adams St Seward Community Jail, Seward AK</u> ✓		
B	Alco S/N <u>X124662</u> ✓ Target Value <u>.079</u> ✓ High Pressure <u>500</u>		
	Alco Test Values 1 st Alco <u>.078</u> ✓ 2 nd Alco <u>.077</u> ✓		
Signature <u>Matthew E. Armstrong</u>		<u>COB</u> <u>2/17/09</u>	
(OVER)			

(Do Not write in the area below)


Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Matthew E Armstrong, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed



Jeanne Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

02/18/09

Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130132

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130132

FEBRUARY 01, 2009 ✓

OPERATOR'S NAME:
ARMSTRONG, MATTHEW E.
OPERATOR'S NUMBER: 5906
SUBJECT'S LAST NAME:
UOC

SUBJECT'S FIRST NAME/MI :
TEST

O.L. #: 0000000
DEPT/AGENCY: SMD1
CASE/REPORT: 09-0000
TEST TYPE: U
ALCO TARGET VALUE: .079
ALCO S/N: X124662 ✓

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.49 in
ALCO TARGET .077 05:36
BLANK TEST .000 05:37
INTERNAL STANDARD VERIFIED 05:37
ALCO TV 29.49 in .078 05:37
BLANK TEST .000 05:38
SUBJECT SAMPLE .000 05:38
BLANK TEST .000 05:39
ALCO TV 29.49 in .077 05:39
BLANK TEST .000 05:40

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130132

FEBRUARY 01, 2009 ✓
TIME 05:33

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 03/29/00

HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 45c

BAROMETER: 29.49 in

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z I \] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z () * #