

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUL 16 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130132</u> ✓	
Name <u>Matthew E. Armstrong</u>		ID# <u>5906</u> ✓	Date <u>22 JUN 10</u> ✓
A	Agency <u>Seward PD</u>	Phone # <u>(907) 224-3338</u>	
Instrument Location <u>410 Adams St Seward Community Jail, Seward AK</u>			
B	Alco S/N <u>633421</u> ✓	Target Value <u>.082</u> ✓	High Pressure <u>550</u>
Alco Test Values		<u>.081</u> ✓ 1 st Alco	<u>.080</u> ✓ 2 nd Alco
Signature <u>Matthew E. Armstrong</u>			<u>COB</u> <u>7/19/10</u>
(OVER)			

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 28th day of July, 2010.

Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130132

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130132

JUNE 22, 2010

OPERATOR'S NAME:

ARMSTRONG, MATTHEW E

OPERATOR'S NUMBER: 5986

SUBJECT'S LAST NAME:

UC

SUBJECT'S FIRST NAME/MI :

TEST

O.I.L. #: A000000

DEPT AGENCY: SMD1

CASE REPORT: 18-0000

TEST TYPE: U

ALCO TARGET VALUE: .082

ALCO S/N: 633421

--- BREATH ANALYSIS ---

.082 ADJUSTED FOR 30.04 in
ALCO TARGET .082 06:00
BLANK TEST .000 06:01
INTERNAL STANDARD *SM* .000 06:01
ALCO TV 30.06 in .081 06:02
BLANK TEST .000 06:02
SUBJECT SAMPLE *CS* .000 06:03
BLANK TEST *7/19/10* .000 06:04
ALCO TV 30.04 in .080 06:04
BLANK TEST .000 06:05

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130132

JUNE 22, 2010

TIME 05:56

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 03/29/00

HEATERS

SAMPLE CHAMBER: 48c

BREATH TUBE: 45c

BAROMETER: *72* 30.04 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
MNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxyz{|}~@