

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

JUL 09 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130132</u>
Name <u>Matthew E. Armstrong</u>	ID# <u>5906</u>	Date <u>3 July 2009</u>
A Agency <u>Seward PD</u>	Phone # <u>(907) 224-3338</u>	<u>NTB</u> 7/16/09
Instrument Location <u>410 Adams St Seward Community Jail, Seward AK</u>		
B Alco SN <u>X301605</u> Target Value <u>.082</u> High Pressure <u>700</u>		
Alco Test Values	<u>.080</u> 1 <sup>st</sup> Alco	<u>.079</u> 2 <sup>nd</sup> Alco
Signature <u>Matthew E. Armstrong</u>		<u>CB</u> 7/14/09


(OVER)

i, Nita J. Bolz, after being first duly sworn, depose and state as follows:  
(Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 10<sup>th</sup> day of July 2009.  
Carolyn M. Noland (Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130132 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130132

JULY 03, 2009

OS 7/14/09

OPERATOR'S NAME:

ORRINGTON, MATTHEW E

OPERATOR'S NUMBER: 5906

SUBJECT'S LAST NAME:

UC

SUBJECT'S FIRST NAME/MI: (94)

UC TEST

O.L. #: 0000000

DEPT/AGENCY: SMD1

CASE/REPORT: 09-0000

TEST TYPE: U

ALCO TARGET VALUE: .002

ALCO S/N: X301605 ✓

### — BREATH ANALYSIS —

.002 ADJUSTED FOR 30.40 in		
ALCO TARGET	.003	10:28
BLANK TEST	.000	10:28
INTERNAL STANDARD	UNCALIB	10:29
ALCO TO 30.40 in	.000	10:29
BLANK TEST	.000	10:30
SUBJECT SAMPLE	.000	10:30
OS 7/14/09		
ALCO TEST	.000	10:31
ALCO TO 30.40 in	.079	10:31
BLANK TEST	.000	10:32

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130132

JULY 03, 2009 ✓

TIME 10:25

### — DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: (94) 03/29/00

NUMBERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 45c

BAROMETER: 30.40 in

FLOW DETECTOR: OKAY

PUMP

MIN SPEED: OKAY

INJECTOR

FILTERS:

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

### PRINTER TEST

!"@#\$(%)&\*+,-./0123456789:;<=>?@BCDEF6  
HIJKLMNPOQRSTUVWXYZ[\]^\_`abcdefgijklmnop  
qrstuvwxyz{|}~