

SEP 09 2010

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm Breath Test Instrument  
State of Alaska*

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

DataMaster cdm S/N <u>130132</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>Matthew E Armstrong</u>	ID# <u>5906</u> Date <u>September 5, 2010</u>
A Agency <u>Seward Police Department</u> Phone # <u>(907) 224-3338</u>	
Instrument Location <u>410 Adams St Seward Community Jail, Seward AK</u>	
B Alco S/N <u>G30459</u> ✓ Target Value <u>.078</u> ✓ High Pressure <u>950</u>	
Alco Test Values <u>.076</u> ✓ 1 <sup>st</sup> Alco <u>.076</u> ✓ 2 <sup>nd</sup> Alco	
Signature <u>Matthew E Armstrong</u>	
COB 9/10/10	
(OVER)	

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
 Nita J. Bolz  
 Scientific Director  
 State Breath Alcohol Program

Subscribed and sworn before me this 9th day of Nov, 2010.

Daniel L. Nicholas (Notary Seal Stamp)  
 Daniel L. Nicholas  
 Notary Public, State of Alaska  
 Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument  
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130132

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130132

SEPTEMBER 05, 2010

OPERATOR'S NAME:  
ARMSTRONG, MATTHEW E  
OPERATOR'S NUMBER: 5906  
SUBJECT'S LAST NAME:  
VOC  
SUBJECT'S FIRST NAME/MI :  
TEST

O.L. #: 0000000  
DEPT/AGENCY: SMD1  
CASE/REPORT: 10-0000  
TEST TYPE: U  
ALCO TARGET VALUE: .078  
ALCO S/N: 630459

### — BREATH ANALYSIS —

.078 ADJUSTED FOR 29.86 in  
ALCO TARGET .077 14:55  
BLANK TEST .000 14:56  
INTERNAL STANDARD VERIFIED 14:56  
ALCO TV 29.86 in .076 14:56  
BLANK TEST .000 14:57  
SUBJECT SAMPLE .000 14:58  
BLANK TEST .000 14:58  
ALCO TV 29.86 in .076 14:59  
BLANK TEST .000 15:00

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130132

SEPTEMBER 05, 2010  
TIME 15:00

### — DIAGNOSTIC CHECK —

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 03/29/00  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 44c  
BAROMETER: 29.86 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[\]^\_`abcde fghi.jklmno  
pqrs tuvwxyz{|}~||