

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130132</u> ✓	
Name	<u>Matthew E. Armstrong</u>	ID#	<u>5906</u> Date <u>25 SEP 08</u> ✓ <u>24 Sep 08</u> (244) 921-08
A Agency	<u>Seward Police Department</u>	Phone #	<u>(907) 224-3338</u>
Instrument Location	<u>Seward Community Jail, Seward AK</u> ✓		
B Alco S/N	<u>X124662</u> ✓	Target Value	<u>.079</u> ✓ High Pressure <u>950</u>
Alco Test Values	<u>.079</u> 1 <sup>st</sup> Alco	<u>.078</u> 2 <sup>nd</sup> Alco	
Signature	<u>[Signature]</u> ✓		<u>CSB</u> <u>9/30/08</u>

(OVER)


Under the Alaska Rules of Evidence, I certify that, (Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Matthew E Armstrong, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



[Signature]

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

10/5/08

Date

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130132 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130132

SEPTEMBER 25, 2008 ✓ *COB 9/24/08*

OPERATOR'S NAME:

ARMSTRONG, MATTHEW E

OPERATOR'S NUMBER: 5906

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

TEST

O.L. #: 1111111

DEPT/AGENCY: SUD1

CASE/REPORT: 00-0000

TEST TYPE: T

ALCO TARGET VALUE: .079 ✓

ALCO S/N: X124662 ✓

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.98 in		
ALCO TARGET	.079	22:13
BLANK TEST	.000	22:14
INTERNAL STANDARD	VERIFIED	22:14
ALCO TV 29.98 in	.079 ✓	22:14
BLANK TEST	.000	22:15
SUBJECT SAMPLE	.000 <i>COB 9/24/08</i>	22:16
BLANK TEST	.000	22:16
ALCO TV 29.98 in	.078 ✓	22:17
BLANK TEST	.000	22:18

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130132

SEPTEMBER 25, 2008 ✓  
TIME 22:08

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	03/29/00
HEATERS	
SAMPLE CHAMBER:	48c
BREATH TUBE:	44c
BAROMETER:	29.99 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST  
!\*"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJKL  
MNPQRSTUVWXYZ[\]^\_`abcdefghijklmnopqrstuvwxyz{|}~■