


VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

OCT 01 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130132 ✓
Name	Doreen R. Valadez	ID#	3423
Date	9/28/09	Agency	Seward PD
Phone #	204-3338	Instrument Location	Seward PD Squad Room ✓
Alco S/N	58442	Target Value	.078
High Pressure	900	Alco Test Value Average	.082
1 st Alco	.082	2 nd Alco	.082
Signature			


CBS
10/12/09

(OVER)


(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 29 day of October, 2009.


(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130132 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130132

SEPTEMBER 28, 2009

OPERATOR'S NAME:

VALADEZ D R

OPERATOR'S NUMBER: 3423

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

X

O.L. #: 0

DEPT/AGENCY: SMD1

CASE/REPORT: 0

TEST TYPE: U

ALCO TARGET VALUE: .078

ALCO S/N: 58442

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 29.91 in
ALCO TARGET .077 16:38
BLANK TEST .000 16:39
INTERNAL STANDARD VERIFIED 16:39
ALCO TV 29.93 in .082 16:40
BLANK TEST .000 16:41
SUBJECT SAMPLE .000 16:41
BLANK TEST .000 16:42
ALCO TV 29.93 in .082 16:43
BLANK TEST .000 16:44

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130132

SEPTEMBER 28, 2009
TIME 16:44

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 03/29/00

HEATERS

SAMPLE CHAMBER: 43c

BREATH TUBE: 43c

BAROMETER: 29.93 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! "# \$ % ' () * + , - . / : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~