

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

DEC 29 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: 130132
13032

Name Doreen R. Valadez ID# 3423 Date 12/21/09

A Agency Seward PD Phone # 224-3332

Instrument Location Seward Community Jail

B Alco S/N 58442 Target Value .078 High Pressure 600

Alco Test Value Average .078 ✓ .079 ✓
1st Alco 2nd Alco

Signature Doreen R. Valadez ✓

(OVER) fmb
1/5/10


(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 14th day of January 2010
Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska


Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130132 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST



ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130132

DECEMBER 21, 2009

OPERATOR'S NAME: *BMB 1/5/10*
VALADEZ D R
OPERATOR'S NUMBER: 3423
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :

X
O.L. #: 0
DEPT/AGENCY: SWD1
CASE/REPORT: 0
TEST TYPE: U
ALCO TARGET VALUE: .078
ALCO S/N: 50442


DAV

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 29.80 in		
ALCO TARGET	.077	16:23
BLANK TEST	.000	16:25
INTERNAL STANDARD	VERIFIED	16:25
ALCO TU 29.81 in	.078	16:25
BLANK TEST	.000	16:26
SUBJECT SAMPLE <i>BMB 1/5/10</i>	.000	16:26
BLANK TEST	.000	16:27
ALCO TU 29.81 in	.079	16:27
BLANK TEST	.000	16:28



DIAGNOSTIC CHECK



ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130132

DECEMBER 21, 2009
TIME 16:29

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	03/29/00
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	44c
BAROMETER:	29.80 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

DAV

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKL MNOPQRSTU VWXYZ[\]^_`abcde fghijklmno
pqrstu vwxyz{|}~