

JAN 06 2011
2011
3/9/11

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130132</u> ✓
Name <u>Matthew E Armstrong</u>		ID# <u>5906</u> ✓	Date <u>January 4, 2011</u> ✓
A	Agency <u>Seward Police Department</u>	Phone #	<u>(907) 224-3338</u>
Instrument Location <u>410 Adams St Seward Community Jail, Seward AK</u>			
Alco S/N <u>X301596</u> ✓		Target Value <u>.076</u> ✓	High Pressure <u>1050</u>
B	Alco Test Values	<u>.071</u> ✓ 1st Alco	<u>.072</u> ✓ 2nd Alco
Signature <u><i>Matthew E Armstrong</i></u>			

(OVER)

AMB
2/8/11

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 10th day of MAR, 2011

Daniel L. Nicholas
 (Notary Seal Stamp)
 Daniel L. Nicholas
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130132

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130132

JANUARY 04, 2011

OPERATOR'S NAME:

ARMSTRONG, MATTHEW E

OPERATOR'S NUMBER: 5906

SUBJECT'S LAST NAME:

UIC

SUBJECT'S FIRST NAME/MI :

TEST

O.L. #: 000000

DEPT/AGENCY: SMD1

CASE/REPORT: 11-0000

TEST TYPE: V

ALCO TARGET VALUE: .076

ALCO S/N: X301596

— BREATH ANALYSIS —

.076 ADJUSTED FOR 29.08 in		
ALCO TARGET	.073	03:39
BLANK TEST	.000	03:40
INTERNAL STANDARD	VERIFIED	03:40
ALCO TU 29.10 in	.071	03:41
BLANK TEST	.000	03:41
SUBJECT SAMPLE	.000	03:42
BLANK TEST	.000	03:43
ALCO TU 29.10 in	.072	03:43
BLANK TEST	.000	03:44

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130132

JANUARY 04, 2011

TIME 03:37

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 03/29/00

HEATERS
SAMPLE CHAMBER: 49c

BREATH TUBE: 44c

BAROMETER: 29.08 in

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxy{|}~