

MAR 0 1 2011 <sup>2011</sup> <sub>AN</sub>

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm Breath Test Instrument  
State of Alaska*

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130132</u>
Name		<u>Matthew E Armstrong</u>	ID# <u>5906</u> Date <u>February 26, 2011</u>
A	Agency	<u>Seward Police Department</u>	Phone # <u>(907) 224-3338</u>
Instrument Location <u>410 Adams St Seward Community Jail, Seward AK</u>			
Alco S/N		<u>X301596</u>	Target Value <u>.076</u> High Pressure <u>1000</u>
Alco Test Values		<u>.075</u> 1st Alco	<u>.074</u> 2 <sup>nd</sup> Alco
Signature		<u><i>Matthew E Armstrong</i></u>	

(OVER)

*BMB*  
*3/6/11*

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

*Nita J. Bolz*  
 Nita J. Bolz  
 Scientific Director  
 State Breath Alcohol Program

Subscribed and sworn before me this 18<sup>th</sup> day of MAR, 2011 <sub>AN</sub>

*[Signature]* (Notary Seal Stamp)

Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument  
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130132

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130132

FEBRUARY 26, 2011

OPERATOR'S NAME:  
ARMSTRONG, MATTHEW E  
OPERATOR'S NUMBER: 5906  
SUBJECT'S LAST NAME:  
UIC

SUBJECT'S FIRST NAME/MI :  
TEST

O.L. #: 0000000  
DEPT/AGENCY: SMD1  
CASE/REPORT: 11-0000  
TEST TYPE: U  
ALCO TARGET VALUE: .076  
ALCO S/N: X301596

### — BREATH ANALYSIS —

.076 ADJUSTED FOR 30.24 in  
ALCO TARGET .076 07:02  
BLANK TEST .000 07:03  
INTERNAL STANDARD VERIFIED 07:03  
ALCO TV 30.25 in .075 07:04  
BLANK TEST .000 07:05  
SUBJECT SAMPLE .000 07:05  
BLANK TEST .000 07:06  
ALCO TV 30.24 in .074 07:06  
BLANK TEST .000 07:07

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130132

FEBRUARY 26, 2011  
TIME 06:59

### — DIAGNOSTIC CHECK —

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE = 03/29/00  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 44c  
BAROMETER: 30.24 in  
FLOW DETECTOR = OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

! "\$ % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~ ¡ ¢