APR 1 : 2011

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130132

Supervisor/Operator Performing the Verification Procedure:

Name Matthew E Armstrong ID# 5906 Date March 23, 2011

A

Agency S Phone # (907) 224-3338

Instrument Location 410 Adams St Seward Community Jail, Seward AK

B

Alco S/N X301596 Target Value .076 High Pressure 900

Alco Test Values .074 (1st Alco) .072 (2nd Alco)

Signature [signature]

(OVER)

BMB 4/11/11

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.

(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.

(3) I am the Scientific Director of the State Breath Alcohol Program.

(4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.

(5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 27 day of MAY, ~~2010~~. NJB 2011

[signature] (Notary Seal Stamp)
~~Daniel L. Nicholas~~ KYLE K LOPFER
Notary Public, State of Alaska
Commission Expires with Office

STATE OF ALASKA SCIENTIFIC CRIME DETECTION LABORATORY

KYLE KLOPFER NOTARY PUBLIC STATE OF ALASKA

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130132

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130132

MARCH 23, 2011

OPERATOR'S NAME:
ARMSTRONG, MATTHEW
OPERATOR'S NUMBER: 5906
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
TEST
O.L. #: 000000
DEPT/AGENCY: SWD1
CASE/REPORT: 11-0000
TEST TYPE: V
ALCO TARGET VALUE: .076
ALCO S/N: X301596

— BREATH ANALYSIS —

.076 ADJUSTED FOR 29.70 in

ALCO TARGET	.075	08:25
BLANK TEST	.000	08:26
INTERNAL STANDARD	VERIFIED	08:26
ALCO TV 29.70 in	.074	08:26
BLANK TEST	.000	08:27
SUBJECT SAMPLE	.000	08:28
BLANK TEST	.000	08:29
ALCO TV 29.72 in	.072	08:29
BLANK TEST	.000	08:30

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130132

MARCH 23, 2011
TIME 08:30

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	03/29/00
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	46c
BAROMETER:	29.72 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~