

JAN 21 2011
2011
DEN
3/2/11

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130133 ✓

Name Garrett Willis ID# 5560 ✓ Date 1-15-11

A Agency AST Phone # 783-0972

Instrument Location Girdwood

B Alco S/N 633421 ✓ Target Value 8.086 ✓ High Pressure 800

Alco Test Value Average .085 ✓ .085 ✓
1st Alco 2nd Alco

Signature Garrett Willis

(OVER)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 11th day of MAR, 2011

Daniel L. Nicholas (Notary Seal Stamp)

Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130133

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130133

JANUARY 15, 2011

OPERATOR'S NAME:

WILLIS/GARRETT/R

OPERATOR'S NUMBER: 5560

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/MI :

UOC

O.L. #: UOC

DEPT/AGENCY: GDWD

CASE/REPORT: UOC

TEST TYPE: U

ALCO TARGET VALUE: .086

ALCO S/N: 633421

--- BREATH ANALYSIS ---

.086 ADJUSTED FOR 30.42 in

ALCO TARGET	.087	13:58
BLANK TEST	.000	13:58
INTERNAL STANDARD	VERIFIED	13:59
ALCO TV 30.42 in	.085	13:59
BLANK TEST	.000	14:00
SUBJECT SAMPLE	.000	14:00
BLANK TEST	.000	14:01
ALCO TV 30.42 in	.085	14:02
BLANK TEST	.000	14:02

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130133

JANUARY 15, 2011

TIME 13:55

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 48c
BAROMETER: 30.42 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLMNPOQRSTUVWXYZ[]^_`abcde fghijklmno
pqrstuvwxyz{|}~■