

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program


DataMaster cdm S/N <u>130133</u>		
Supervisor/Operator Performing the Verification Procedure:		
Name <u>Peter Steen</u>	ID# <u>4572</u> Date <u>3/13/11</u>	
A Agency <u>Palmer PD AST 1 BHP</u> Phone # <u>373-8342</u>		
Instrument Location <u>Girdwood AST</u>		
B Alco S/N <u>633421</u> Target Value <u>.086</u> High Pressure <u>650</u>		
Alco Test Values	<u>.083</u> 1st Alco	<u>.083</u> 2nd Alco
Signature		
(OVER)		

BMB  
5/9/11

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program



Subscribed and sworn before me this 27 day of MAY, 2011.

  
Daniel L. Nicholas (Notary Seal Stamp)  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130133

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130133

MARCH 13, 2011

OPERATOR'S NAME:  
STEEN, PETER, C  
OPERATOR'S NUMBER: 4572  
SUBJECT'S LAST NAME:  
UOC  
SUBJECT'S FIRST NAME/MI:  
UOC

O.L. #: UOC  
DEPT/AGENCY: PALT  
CASE/REPORT: UOC  
TEST TYPE: V  
ALCO TARGET VALUE: .086  
ALCO S/N: 633421

### — BREATH ANALYSIS —

.086 ADJUSTED FOR 29.90 in  
ALCO TARGET .085 18:35  
BLANK TEST .000 18:36  
INTERNAL STANDARD VERIFIED 18:36  
ALCO TV 29.90 in .083 18:36  
BLANK TEST .000 18:37  
SUBJECT SAMPLE .000 18:38  
BLANK TEST .000 18:38  
ALCO TV 29.90 in .083 18:39  
BLANK TEST .000 18:40

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130133

MARCH 13, 2011  
TIME 18:32

### — DIAGNOSTIC CHECK —

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 42c  
BAROMETER: 29.90 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

! " # \$ % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~ ¡ ¢