

FEB 20 2009

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130133 ✓

Name BRYAN BARLOW ID# 4548 Date 2/19/09 ✓

A Agency ALASKA STATE TROOPERS Phone # 907-783-0972

Instrument Location Girdwood AST

B Alco S/N X104671 ✓ Target Value .077 ✓ High Pressure 550

Alco Test Value Average .081 2/19/09 CMJ .081 2/20/09 CMJ

1st Alco 2nd Alco

Signature Bryan Barlow ✓ CB
3/2/09

Under the Alaska Rules of Evidence, I certify that; (OVER)

(Do Not write in the area below)
(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Bryan W Barlow, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

Jeanne Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

3/2/09

Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130133 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
 DATAMASTER cdm 130133
 FEBRUARY 19, 2009 ✓
 OPERATOR'S NAME: BARLOW BRYAN W
 OPERATOR'S NUMBER: 4540
 SUBJECT'S LAST NAME: _____
 CALIBRATION
 SUBJECT'S FIRST NAME/MI: _____
 VERIFICATION OF
 O.I.L. #: 1234567
 DEPT/AGENCY: 6040
 CASE/REPORT: 09-12345
 TEST TYPE: U
 ALCO TARGET VALUE: .077 ✓
 ALCO S/N: X124671 ✓

--- BREATH ANALYSIS ---

.077 ADJUSTED FOR 30.79 in		
ALCO TARGET	.079	20:36
BLANK TEST	.000	20:37
INTERNAL STANDARD	VERIFIED	20:37
ALCO TV 30.79 in	.081 ✓	20:38
BLANK TEST	.000	20:39
SUBJECT SAMPLE	.000	20:39
BLANK TEST	.000	20:40
ALCO TV 30.79 in	.081 ✓	20:41
BLANK TEST	.000	20:41

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
 DATAMASTER cdm 130133
 FEBRUARY 19, 2009 ✓
 TIME 20:31

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
 PROGRAM: OKAY
 SOFTWARE DATE: 02/20/01
 HEATERS
 SAMPLE CHAMBER: 49c
 BREATH TUBE: 41c
 BAROMETER: 30.79 in
 FLOW DETECTOR: OKAY
 PUMP
 HIGH SPEED: OKAY
 DETECTOR: OKAY
 FILTERS: OKAY
 QUARTZ STANDARD: OKAY
 CALIBRATION: OKAY

PRINTER TEST
 !"#%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
 KLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
 qrstuvwxyz{|}~■