VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130133

Supervisor/Operator Performing the Verification Procedure:

A

Name Bryan Barlow ID# 4548 Date 3/26/08

Agency Alaska State Troopers Phone # 9077830972

Instrument Location Girdwood AST

B

Alco S/N X172942 Target Value .079 High Pressure 500 lbs.

Alco Test Value Average .078 (1st Alco) .079 (2nd Alco)

Signature Bryan Barlow CDB 4/7/08

(OVER)

Under the Alaska Rules of Evidence, I certify that;

(Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Bryan W Barlow, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed

4/12/08

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

Date

APR 01 2008

BT9 03/00

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130133

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130133

MARCH 26, 2008 COB 4/7/08

OPERATOR'S NAME:
BARLOW BRYAN W
OPERATOR'S NUMBER: 4548
SUBJECT'S LAST NAME:
VERIFICATION
SUBJECT'S FIRST NAME/MI :
OF CALIBRATION
O.L. #: 1234567
DEPT/AGENCY: GDWO
CASE/REPORT: 08-12345
TEST TYPE: V
ALCO TARGET VALUE: .079
ALCO S/N: X172942

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 30.30 in

ALCO TARGET	.080	13:35
BLANK TEST	.000	13:36
INTERNAL STANDARD	VERIFIED	13:36
ALCO TV 30.32 in	.078	13:37
BLANK TEST	.000	13:38
SUBJECT SAMPLE	.000 COB 4/7/08	13:38
BLANK TEST	.000	13:39
ALCO TV 30.32 in	.079	13:40
BLANK TEST	.000	13:40

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130133

MARCH 26, 2008
TIME 13:42

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	48c
BREATH TUBE:	41c
BAROMETER:	30.32 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~