

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130133 ✓
A	Name	Bryan Barlow	ID# 4548 Date 5/31/08 ✓
	Agency	Alaska State Troopers ✓	Phone # 907-783-0972
	Instrument Location	Girdwood AST	
B	Alco S/N	X172942 ✓	Target Value .079 ✓ High Pressure 400
	Alco Test Value Average	.078 ✓ 1 <sup>st</sup> Alco	.079 ✓ 2 <sup>nd</sup> Alco
	Signature		 6/5/08

(OVER)

Under the Alaska Rules of Evidence, I certify that;


(Do Not write in the area below)


(a) I am Jeane Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Bryan W Barlow; whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,





Jeane Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

6/9/08  
Date

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130133

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130133

MAY 31, 2008

OPERATOR'S NAME: **CB**  
BARLOW BRYAN W  
OPERATOR'S NUMBER: 4548  
SUBJECT'S LAST NAME: **W/6/108**  
CALIBRATION  
SUBJECT'S FIRST NAME/MI :  
VERIFICATION OF  
D.L.E. #: 1234567  
DEPT/AGENCY: 6D/00  
CASE/REPORT: 08-12345  
TEST TYPE: U  
ALCO TARGET VALUE: .079  
ALCO S/N: X172942

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 30.45 in  
ALCO TARGET .000 19:53  
BLANK TEST .000 19:54  
INTERNAL STANDARD VERIFIED 19:54  
ALCO TV 30.45 in .078 19:55  
BLANK TEST .000 19:55  
SUBJECT SAMPLE .000 **CB** 19:56  
BLANK TEST .000 **CB** 19:57  
ALCO TV 30.45 in .079 19:57  
BLANK TEST .000 19:58

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130133

MAY 31, 2008  
TIME 19:50

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 41c  
BAROMETER: 30.45 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFG  
HIJKLMNPOQRSTUVWXYZ[\]^\_`abcde fghijklmno  
pqrstuvwxyz{|}~■