VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: **DataMaster cdm S/N** 130133 ✓

A

Name BRYAN BARLOW ID# 4548 Date 8/7/08 ✓

Agency Alaska State Troopers Phone # 907-783-0972

Instrument Location Girdwood AST ✓

B

Alco S/N X172942 ✓ Target Value .079 ✓ High Pressure 250

Alco Test Value Average .078 ✓ (1st Alco) .079 ✓ (2nd Alco)

Signature Bryan Barlow ✓

CCB 8/12/08

(OVER)

Under the Alaska Rules of Evidence, I certify that;

(Do Not write in the area below)

(a) I am Jeanne Swartz , Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Bryan W Barlow , whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

STATE OF ALASKA SCIENTIFIC CRIME DETECTION LABORATORY

Signed and sealed,

8/13/08

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

Date

BT9 03/00

AUG 11 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130133 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130133

AUGUST 07, 2008 ✓ CDB 8/12/08

OPERATOR'S NAME:
BARLOW BRYAN W
OPERATOR'S NUMBER: 4548
SUBJECT'S LAST NAME:
VERIFICATION
SUBJECT'S FIRST NAME/MI :
OF CALIBRATION
D.L. #: 1234567
DEPT/AGENCY: GDW0
CASE/REPORT: 08-12345
TEST TYPE: V
ALCO TARGET VALUE: .079 ✓
ALCO S/N: X172942 ✓

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 30.25 in

ALCO TARGET	.079	17:50
BLANK TEST	.000	17:51
INTERNAL STANDARD	VERIFIED	17:51
ALCO TV 30.25 in	.078 ✓	17:52
BLANK TEST	.000	17:52
SUBJECT SAMPLE	.000 CDB 8/12/08	17:53
BLANK TEST	.000	17:54
ALCO TV 30.25 in	.079 ✓	17:54
BLANK TEST	.000	17:55

5/13/08

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130133

AUGUST 07, 2008 ✓
TIME 17:45

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	30.25 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~