

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130133 ✓
Name		Bryan Barlow	ID# 4548 Date 9/20/08 ✓
A	Agency	Alaska State Troopers	Phone # 907-783-0972
Instrument Location		Girdwood AST	✓
Alco S/N		X124671 ✓	Target Value .077 ✓
			High Pressure 800 ✓
Alco Test Value Average		0.081 ^{COB} / 9/20/08 ✓	0.081 ^{COB} / 9/20/08 ✓
		1 st Alco	2 nd Alco
Signature		Bryan Barlow ✓	COB / 9/20/08

(OVER)

Under the Alaska Rules of Evidence, I certify that;

(Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. instrument operator, Bryan W Barlow, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed

Juanne Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

11/21/08
Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130133 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130133

SEPTEMBER 20, 2008 *COB 11/20/08*

OPERATOR'S NAME:
BARLOW BRYAN W
OPERATOR'S NUMBER: 4548
SUBJECT'S LAST NAME:
VERIFICATION
SUBJECT'S FIRST NAME/MI :
OF CALIBRATION
O.L. #: 1234567
DEPT/AGENCY: GD40
CASE/REPORT: 08-12345
TEST TYPE: V
ALCO TARGET VALUE: .077 ✓
ALCO S/N: X124671

--- BREATH ANALYSIS ---

.077 ADJUSTED FOR 30.22 in		
ALCO TARGET	.077	19:07
BLANK TEST	.000	19:08
INTERNAL STANDARD	VERIFIED	19:08
ALCO TV 30.22 in	.081 ✓	19:09
BLANK TEST	.000	19:09
SUBJECT SAMPLE <i>COB 11/20/08</i>	.000	19:10
BLANK TEST	.000	19:11
ALCO TV 30.22 in	.081 ✓	19:11
BLANK TEST	.000	19:12

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130133

SEPTEMBER 20, 2008 ✓
TIME 19:04

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01

HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 42c

BAROMETER: 30.22 in

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST
!*"#&?'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKL MNOPQRSTUVWXYZ[\]^_`abcde fghi jk lmnop
qrstuvwxy z{|}~■