

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAR 23 2010

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130133 ✓

Name BRYAN BARLOW ID# 4548 Date 3/19/10

A Agency AST Phone # 907-783-0972

Instrument Location Girdwood ✓

B Alco S/N 75260 Target Value .079 High Pressure 300

Alco Test Value Average .078 .079 ✓
1st Alco 2nd Alco

Signature [Handwritten Signature] ✓

(OVER)

BMB
3/24/10

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (Do Not write in the area below)
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
 - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
 - (3) I am the Scientific Director of the State Breath Alcohol Program.
 - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
 - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
 - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Handwritten Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 15th day of April, 2010.

[Handwritten Signature] (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster edm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster edm S/N 130133 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER edm 130133 ✓

MARCH 19, 2010 ✓

OPERATOR'S NAME: BARLOW BRYAN W
OPERATOR'S NUMBER: 4548
SUBJECT'S LAST NAME: CALIBRATION
SUBJECT'S FIRST NAME/MI :
VERIFICATION OF
D.L. #: 1234567
DEPT/AGENCY: GDWD
CASE/REPORT: 10-12345 ✓
TEST TYPE: U ✓
ALCO TARGET VALUE: .079 ✓
ALCO S/N: 75260 ✓

BMB
3/24/10

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 30.25 in ✓
ALCO TARGET .079 14:58
BLANK TEST .000 15:00
INTERNAL STANDARD VERIFIED 15:00
ALCO TV 30.25 in .078 ✓ 15:00
BLANK TEST .000 15:01
SUBJECT SAMPLE *BMB* .000 15:02
BLANK TEST .000 15:02
ALCO TV 30.25 in *3/24/10* .079 ✓ 15:03
BLANK TEST .000 15:03

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER edm 130133 ✓

MARCH 19, 2010 ✓
TIME 15:04 ✓

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 39c
BAROMETER: 30.25 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z I \] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ ¡ ¢