

AUG 17 2010

VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm Breath Test Instrument
State of Alaska*

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program


Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130133</u> ✓
Name		<u>BRYAN BARLOW</u>	ID# <u>4548</u> Date <u>8/15/10</u> ✓
A	Agency	<u>Alaska State Troopers</u>	Phone # <u>907-783-0972</u>
Instrument Location		<u>Girdwood AST</u>	
Alco S/N		<u>X172966</u> ✓	Target Value <u>.084</u> ✓ High Pressure <u>500 lbs.</u>
Alco Test Values		<u>.084</u> ✓ 1st Alco	<u>.083</u> ✓ 2 nd Alco
Signature			<u>CUB</u> <u>8/17/10</u>

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 23RD day of SEPT, 2010.


 Daniel L. Nicholas
 Notary Public, State of Alaska
 Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130133

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130133

AUGUST 15, 2010 *CSB 8/15/10*

OPERATOR'S NAME:
BARLOW BRYAN W
OPERATOR'S NUMBER: 4548
SUBJECT'S LAST NAME:
CALIBRATION
SUBJECT'S FIRST NAME/MI :
VERIFICATION OF
O.L. #: 1234567
DEPT/AGENCY: GDWD
CASE/REPORT: 10-123456
TEST TYPE: U
ALCO TARGET VALUE: .004
ALCO S/N: X172966

--- BREATH ANALYSIS ---

.004 ADJUSTED FOR 30.35 in		
ALCO TARGET	.005	19:09
BLANK TEST	.000	19:10
INTERNAL STANDARD	VERIFIED	19:11
ALCO TV 30.35 in	.004	19:11
BLANK TEST	.000	19:12
SUBJECT SAMPLE	.000	19:12
BLANK TEST	.000	19:13
ALCO TV 30.35 in	.003	19:13
BLANK TEST	.000	19:14

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130133

AUGUST 15, 2010
TIME 19:15

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 42c

BAROMETER: 30.35 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!*"#%&'()*+,-./0123456789:;<=>?@ABCDEFGHI
HIJKL MNOPQRSTUVWXYZ\]^_`abcdefgijklmnop
qrsstuvwxyz{|}~