

OCT 26 2010

VERIFICATION OF CALIBRATION REPORT
of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130133 ✓

Name BRYAN BARLOW ID# 4548 ✓ Date 10/15/10 ✓

Agency Alaska State Troopers Phone # 907-783-0972

Instrument Location Gridwood AST

Alco S/N X172966 ✓ Target Value .084 ✓ High Pressure 425

Alco Test Value Average .083 ✓ .084 ✓
 1st Alco 2nd Alco

Signature Bryan Barlow

(OVER)

BMB
12/9/10

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
 Nita J. Bolz
 Scientific Director

Subscribed and sworn before me on the 17th day of DEC, 2010

Daniel L. Nicholas
 Daniel L. Nicholas
 Notary Public, State of Alaska
 My Office is located at: [illegible]



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130133

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

Breath Test Record			
Serial Number	130133	Model	DataMaster
Location	GDW0	AST/Girdwood	
Subject Information			
Last Name	Calibration	O.L. #	1234567
First Name	Verification Of		
Operator Information			
Operator Name	Barlow Bryan W	Case	10-123456
Operator #	4548	Department	GDW0
Test Results			
Alco Target	.084	Alco Serial	X172966
Nominal	0.084	Barometer	30.24
Dry Gas 1	0.083	Time	21:33:59
Subject Alcohol	0.000	Time	21:35:25
Dry Gas 2	0.084	Time	21:36:34
		Test Type	<input checked="" type="checkbox"/>
		Test Date	10/15/2010
		Test Time	21:32:22
		Report	<input checked="" type="checkbox"/>

Diagnostic Test Record			
Serial Number	130133	Model	DataMaster
Location	GDW0	AST/Girdwood	
Diagnostic Test Results			
Test Date	10/15/2010	Computer	Okay
Test Time	21:29:46	Program	Okay
		Software Date	02/20/2001
		Sample Chamber (C)	48
		Breath Tube (C)	42
		Barometer (in)	30.25
		Flow Detector	Okay
		High Speed Pump	Okay
		Detector	Okay
		Filter	Okay
		Quartz	Okay
		Calibration	Okay

