

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130133

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130133

DECEMBER 14, 2010 *COB 12/30/10*

OPERATOR'S NAME: WILLIS/GARRETT/R
OPERATOR'S NUMBER: 5560
SUBJECT'S LAST NAME: UOC
SUBJECT'S FIRST NAME/MI: UOC
O.L. #: UOC
DEPT/AGENCY: GDWD *COB*
CASE/REPORT: UOC
TEST TYPE: U
ALCO TARGET VALUE: .004
ALCO S/N: X172966

--- BREATH ANALYSIS ---

| | | |
|------------------------------------|----------|-------|
| .004 ADJUSTED FOR 29.67 in | | |
| ALCO TARGET | .003 | 14:59 |
| BLANK TEST | .000 | 15:01 |
| INTERNAL STANDARD | VERIFIED | 15:01 |
| ALCO TV 29.67 in | .002 | 15:01 |
| BLANK TEST | .000 | 15:02 |
| SUBJECT SAMPLE <i>COB 12/30/10</i> | .000 | 15:02 |
| BLANK TEST | .000 | 15:03 |
| ALCO TV 29.67 in | .002 | 15:04 |
| BLANK TEST | .000 | 15:04 |

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130133

DECEMBER 14, 2010
TIME 14:46

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY *COB*
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.65 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!""#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHI
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnop
qrstuvwxyz{|}~■