

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130134</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>Tyron Walker</u>	ID# <u>4555</u> Date <u>4/15/11</u>
A Agency <u>Ketchikan Police Dept.</u>	Phone # <u>225-6631</u>
Instrument Location <u>361 Main St Ktn AK 99901</u>	
B Alco S/N <u>70729</u> Target Value <u>.084</u> High Pressure <u>400</u>	
Alco Test Values	<u>.087</u> <u>.087</u>
1st Alco	2 nd Alco
Signature <u>Tyron Walker</u>	
(OVER)	

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 31 day of MAY, 2010. NJB 2011

Kyle Klopfer (Notary Seal Stamp)
~~Daniel L. Nicholas~~ KYLE KLOPFER
Notary Public, State of Alaska
Commission Expires with Office



CRF
4/21/11

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130134

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130134

APRIL 15, 2011

CTF 4/21/11

OPERATOR'S NAME:

WALKER TV

OPERATOR'S NUMBER: 4555

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/NI :

U

O.L. #: 12345

DEPT/AGENCY: KET1

CASE/REPORT: 11-1234

TEST TYPE: U

ALCO TARGET VALUE: .004

ALCO S/N: 70729

--- BREATH ANALYSIS ---

.004 ADJUSTED FOR 30.30 in		
ALCO TARGET	.005	13:19
BLANK TEST	.000	13:20
INTERNAL STANDARD	VERIFIED	13:20
ALCO TV 30.30 in	.007	13:20
BLANK TEST	.000	13:21
SUBJECT SAMPLE	.000	13:22
BLANK TEST	.000	13:23
ALCO TV 30.30 in	.007	13:23
BLANK TEST	.000	13:24

CTF
4/21/11

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130134

APRIL 15, 2011

TIME 13:25

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	30.30 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKLNPQRSTUWXYZ[\]^_`abcde fghi jk lmnop
qrs tuvwxyz{|}~