

FEB 08 2008 VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: **DataMaster cdm S/N** 130134

Name HUTT, William T. ID# 3359 Date 01/31/08

Agency AST / Anchor Point Post (ANPE) Phone # 235-8239
235-3000

Instrument Location 34115 Sterling Hwy. / P.O. Box 807 Anchor Point,
AK 99556

Alco S/N X301536^v Target Value .075^v High Pressure 450 psi

Alco Test Value Average .076 .077
 1st Alco 2nd Alco

Signature William T. Hutt CSB
2/2/08

Under the Alaska Rules of Evidence, I certify that (OVER)

- (a) I am Jeanne Swartz, Criminalist (ID Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.
- (b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, William Hutt, whose name appears in my records, is qualified to verify the calibration of the above instrument.
- (c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed

Jeanne Swartz 02/22/08
 Jeanne Swartz
 Scientific Director Designee
 State of Alaska
 State Breath Alcohol Testing Program
 Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130134

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130134

JANUARY 31, 2008

OPERATOR'S NAME: *WTH*
HUTT, WILLIAM T
OPERATOR'S NUMBER: 3359
SUBJECT'S LAST NAME: *Sample*
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
ETOH

O.L. #: 01234567
DEPT/AGENCY: AST0 / ANPE
CASE/REPORT: 08-00000
TEST TYPE: U
ALCO TARGET VALUE: .075
ALCO S/N: X301536

--- BREATH ANALYSIS ---

01/31/08
0.075 ADJUSTED FOR 29.72 in
ALCO TARGET .074 01:45
BLANK TEST .000 01:46
INTERNAL STANDARD VERIFIED 01:46
ALCO TV 29.72 in .076 01:46
BLANK TEST .000 01:47
SUBJECT SAMPLE *WTH* .000 01:48
BLANK TEST *01/31/08* .000 01:49
ALCO TV 29.72 in .077 *01/31/08* 01:49
BLANK TEST .000 01:50

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130134

JANUARY 31, 2008
TIME 01:40

WTH

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 29.72 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKL MNOPQRSTUVWXYZ \] ^ _ ` abcdefghijklmno
pqrstuvwxyz{|}~