VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130134

Supervisor/Operator Performing the Verification Procedure:

A

Name HUTT, William T. ID# 3359 Date 03/27/08

Agency AST / Anchor Point Post (ANPE) Phone # 235-8239 235-3000

Instrument Location 34115 Sterling Hwy. / P.O. Box 807 Anchor Point, AK 99603

B

Alco S/N X301536 Target Value .075 High Pressure 450 psi

Alco Test Value Average .078 (1st Alco) .078 (2nd Alco)

Signature William T. Hutt

4/7/08

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz , Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, William Hutt , whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

4/12/08

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

Date

BT9 03/00

APR 01 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130134

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130134

MARCH 27, 2008 CGB 4/4/08

OPERATOR'S NAME:
HUTT, WILLIAM T
OPERATOR'S NUMBER: 3359
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
ETOH
O.L. #: 01234567
DEPT/AGENCY: AST0 / AMPE
CASE/REPORT: 08-00000
TEST TYPE: V
ALCO TARGET VALUE: .075
ALCO S/N: X301536

--- BREATH ANALYSIS ---

.075 ADJUSTED FOR 30.12 in

ALCO TARGET	.075	23:35
BLANK TEST	.000	23:35
INTERNAL STANDARD	VERIFIED	23:36
ALCO TV 30.12 in	.078	23:36
BLANK TEST	.000	23:37
SUBJECT SAMPLE	.000	23:38
BLANK TEST	.000	23:38
ALCO TV 30.12 in	.078	23:39
BLANK TEST	.000	23:39

WTHD 03/27/08 CGB 4/4/08

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130134

MARCH 27, 2008 WTHD
TIME 23:31

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	30.14 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~