

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

AUG 20 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130134</u> ✓
Name <u>J Swartz</u>		ID# <u>4314</u> Date <u>8/17/09</u> ✓
A	Agency <u>SCDL</u>	Phone # <u>269-5592</u>
Instrument Location <u>SCDL</u> ✓		
Alco S/N <u>SCOTT08</u> ✓		Target Value <u>.080</u> ✓ High Pressure <u>600 psi</u>
B	Alco Test Values	
	<u>.080</u> ✓ 1 st Alco	<u>.079</u> ✓ 2 nd Alco
Signature <u>[Signature]</u> ✓		<u>CSB</u> <u>8/31/09</u>
(OVER)		

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 29th day of Sept, 2009.

[Signature] (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



