

JUN 21 2010

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: **DataMaster cdm S/N** 130134 ✓

Name Derek A McGarrigan ID# 5420 ✓ Date 6/16/10 ✓

A Agency Ketchikan Police Department Phone # 225-6631

Instrument Location Ketch. Police Dept, 361 Main St, Ketch. AK 99901

---

B Alco S/N X301765 Target Value .079 ✓ High Pressure 900

Alco Test Values .081 ✓ .082 ✓  
 1<sup>st</sup> Alco 2<sup>nd</sup> Alco

Signature [Handwritten Signature]

(OVER)

BMB  
6/23/10

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Handwritten Signature: Nita J. Bolz]  
 Nita J. Bolz  
 Scientific Director  
 State Breath Alcohol Program

Subscribed and sworn before me this 23<sup>rd</sup> day of July, 2010.

[Handwritten Signature: Carolyn M. Noland]  
 Carolyn M. Noland  
 Notary Public, State of Alaska  
 Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130134

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130134

JUNE 16, 2010

OPERATOR'S NAME: *BMMB 6/23/10*  
MCBARRIGAN DEREK A  
OPERATOR'S NUMBER: 5428  
SUBJECT'S LAST NAME:  
UOC

SUBJECT'S FIRST NAME/MI :  
UOC

O.L. #: 1234567  
DEPT/AGENCY: KET1  
CASE/REPORT: 10-1  
TEST TYPE: U  
ALCO TARGET VALUE: .079  
ALCO S/N: X301765

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 30.33 in  
ALCO TARGET .000 22:40  
BLANK TEST .000 22:41  
INTERNAL STANDARD VERIFIED 22:41  
ALCO TV 30.33 in .001 22:41  
BLANK TEST .000 22:42  
SUBJECT SAMPLE *BMMB 6/23/10* .000 22:43  
BLANK TEST .000 22:43  
ALCO TV 30.33 in .002 22:43  
BLANK TEST .000 22:44

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130134

JUNE 16, 2010  
TIME 22:37

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 01/08/09  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 41c  
BAROMETER: 30.33 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnop  
qrstuvwxyz{|}~■