



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130135

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130135

MARCH 20, 2012

OPERATOR'S NAME: SANTORA JEANNINE L  
OPERATOR'S NUMBER: 2276  
SUBJECT'S LAST NAME: VERIFICATION  
SUBJECT'S FIRST NAME/NI :  
OF CAL  
O.L. #: 0000000  
DEPT/AGENCY: GALO  
CASE/REPORT:  
TEST TYPE: V  
ALCO TARGET VALUE: .081  
ALCO S/N: 78855

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 30.16 in  
ALCO TARGET .081 10:01  
BLANK TEST .000 10:02  
INTERNAL STANDARD VERIFIED 10:02  
ALCO TO 30.16 in .082 10:03  
BLANK TEST .000 10:03  
SUBJECT SAMPLE .000 10:04  
BLANK TEST .000 10:05  
ALCO TO 30.17 in .081 10:05  
BLANK TEST .000 10:06

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130135

MARCH 20, 2012  
TIME 10:10

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 41c  
BAROMETER: 30.17 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

\*#2K'()++-./0123456789:;<=>@ABCDEFG  
HIJKLNPQRSTUWXYZ[]^\_`abcde fghijklmno  
pqrstuvwxyz{|}~