

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N: 130135

Supervisor/Operator Performing the Verification Procedure:

Name: Trp. J. Santora

ID# 2276

Date: 5/9/12 ~~5-8-12~~ BMB 5/17/12

A

Agency: Alaska State Troopers

Phone # 656-1233

Instrument Location: Galena Police Department

B

Alco S/N: 78855

Target Value: .081

High Pressure: 600

Alco Test Values

.080

.080

1<sup>st</sup> Alco

2<sup>nd</sup> Alco

Signature

J. Santora  
(OVER)

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 3rd day of July, ~~2010~~ 2012 NAR 7/3/12

Nikki Roth  
Nikki Roth, Notary Public  
My Commission Expires With Office



BMB  
5/17/12

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130135

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130135

MAY 09, 2012

OPERATOR'S NAME: *Bomb 5/17/12*  
SANTORA JEANNINE L  
OPERATOR'S NUMBER: 2276  
SUBJECT'S LAST NAME:  
VERIFICATION  
SUBJECT'S FIRST NAME/NI :  
OF CAL  
O.L. #: 0000000  
DEPT/AGENCY: GALO  
CASE/REPORT:  
TEST TYPE: V  
ALCO TARGET VALUE: .081  
ALCO S/N: 78855

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 29.70 in  
ALCO TARGET .080 13:44  
BLANK TEST .000 13:45  
INTERNAL STANDARD VERIFIED 13:46  
ALCO TU 29.70 in .080 13:46  
BLANK TEST .000 13:47  
SUBJECT SAMPLE .000 13:48  
BLANK TEST .000 13:49  
ALCO TU 29.70 in .080 13:49  
BLANK TEST .000 13:50

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130135

MAY 09, 2012  
TIME 13:51

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 43c  
BAROMETER: 29.70 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@BCDEFG  
HIJKL/MNOPQRSTU/VWXYZ\]^\_`abcde fghijklano  
pqrs tuvwxyz{|}~