

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

MAR 16 2009

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130135</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>Michael Potter (AST)</u> ID# <u>5841</u> Date <u>3/13/09</u> ✓	
A	Agency <u>Alaska State Troopers</u> Phone # <u>656-1634</u>
Instrument Location <u>Calena P.D.</u> ✓	
Alco S/N <u>78259</u> ✓ Target Value <u>.083</u> ✓ High Pressure <u>15. psi</u>	
B	Alco Test Value Average <u>.080</u> ✓ <u>.080</u> ✓ 1 st Alco 2 nd Alco
Signature <u>[Signature]</u> ✓ COB 3/24/09	
(OVER)	

Under the Alaska Rules of Evidence, I certify that: (Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Michael Potter, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

[Signature] 3/24/09

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130135 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130135

MARCH 13, 2009 ✓ *COB 3/24/09*

OPERATOR'S NAME: POTTER/MICHAEL/J
OPERATOR'S NUMBER: 5841
SUBJECT'S LAST NAME: UDC
SUBJECT'S FIRST NAME/MI: UDC
IDCL #: UDC
DEPT/AGENCY: D
CASE REPORT: X
TEST TYPE: U
ALCO TARGET VALUE: .083 ✓
ALCO S/N: 78259 ✓

--- BREATH ANALYSIS ---

.083 (ADJUSTED FOR 29.91 in		
ALCO TARGET	.083	10:36
BLANK TEST	.000	10:38
INTERNAL STANDARD	VERIFIED	10:38
ALCO TU 29.91 in	.080 ✓	10:38
BLANK TEST	.000	10:39
SUBJECT SAMPLE <i>COB 3/24/09</i>	.000	10:39
BLANK TEST	.000	10:40
ALCO TU 29.91 in	.080 ✓	10:41
BLANK TEST	.000	10:41

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130135

MARCH 13, 2009 ✓
TIME 10:43

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01

HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 43c

BAROMETER: 29.91 in

FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST
!@#%&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKL MNOPQRSTU VWXYZ[\]^_`a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~