VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130135

Supervisor/Operator Performing the Verification Procedure:

A

Name JSwartz ID# 4314 Date 10/21/08

Agency SCDL Phone # 269-5592

Instrument Location SCDL

B

Alco S/N AG 4098 Target Value .080 High Pressure 500 psi

Alco Test Values .077 (1st Alco) .078 (2nd Alco)

Signature [signature] CQB 11/20/08

(OVER)

(Do not write in the area below)

Under the Alaska Rules of Evidence, I certify that:

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Jeanne M Swartz, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed

[signature] 11/21/08

Jeanne Swartz — Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

NOV 20 2008

BT9 06/03

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130135

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

Breath Test Record

Serial Number: 130135 Model: DataMaster
Location: TOG1 Togiak Police Dept.

CCB 11/20/08

Subject Information

Last Name: Verification Of Cal O.L. #: L
First Name: L

Operator Information

Operator Name: Swartz/J Case:
Operator #: 4314 Department: AAA1

Test Results

Alco Target: .080 Alco Serial: AG4098
Nominal: 0.079 Barometer: 29.60 Test Type: V
Dry Gas 1: 0.077 Time: 18:22:15 Test Date: 10/21/2008
Subject Alcohol: 0.000 Time: 18:23:40 Test Time: 18:20:55
Dry Gas 2: 0.078 Time: 18:24:43 Report: Y

CCB 11/20/08

10/21/08

DIAGNOSTIC CHECK

Diagnostic Test Record

Serial Number: 130135 Model: DataMaster
Location: TOG1 Togiak Police Dept.

Diagnostic Test Results

Test Date: 10/21/2008
Test Time: 18:13:27

Computer: Okay
Program: Okay
Software Date: 02/20/2001
Sample Chamber (C): 48
Breath Tube (C): 41
Barometer (in): 29.60
Flow Detector: Okay
High Speed Pump: Okay
Detector: Okay
Filter: Okay
Quartz: Okay
Calibration: Okay