

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

JUL 29 2010

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130135</u> ✓	
Supervisor/Operator Performing the Verification Procedure:	
Name	<u>Trooper Jeannine Santora</u> ID# <u>2276</u> ✓ Date <u>7/23/10</u> ✓
A Agency	<u>Alaska State Troopers</u> Phone # <u>907 656-1233</u>
Instrument Location <u>Galena Police Department</u>	
Alco S/N <u>78855</u> ✓ Target Value <u>.081</u> High Pressure <u>1150 psi</u>	
B Alco Test Value Average	<u>.080</u> ✓ <u>.079</u> ✓ 1 st Alco 2 nd Alco
Signature	<u>[Signature]</u> <u>COB</u> <u>7/30/10</u>
(OVER)	

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 9th day of AUGUST, 2010.

[Signature] (Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130135

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130135

JULY 23, 2010

CBS
7/30/10

OPERATOR'S NAME:
SANTORA, JEANNINE/L
OPERATOR'S NUMBER: 2276
SUBJECT'S LAST NAME:
VERIFICATION
SUBJECT'S FIRST NAME/MI :
OF CAL
O.L. #: -
DEPT/AGENCY: GALO
CASE/REPORT:
TEST TYPE: V
ALCO TARGET VALUE: .081
ALCO S/N: 78855

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 29.88 in		
ALCO TARGET	.088	08:45
BLANK TEST	.000	08:46
INTERNAL STANDARD	VERIFIED	08:46
ALCO TO 29.88 in	.000	08:47
BLANK TEST	.000	08:47
SUBJECT SAMPLE	.000	08:48
BLANK TEST	.000	08:49
ALCO TO 29.88 in	.079	08:49
BLANK TEST	.000	08:50

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130135

JULY 23, 2010
TIME 08:51

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/28/01
HEATERS	
SAMPLE CHAMBER:	48c
BREATH TUBE:	42c
BAROMETER:	29.88 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()*+,-./:0123456789:;<=>?@BCDEFG
HIJKL/MNOPQRSTU/VWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~#