

VERIFICATION OF CALIBRATION REPORT JAN 1 2011
of DataMaster cdm Breath Test Instrument
 State of Alaska
 Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N: 130135 ✓

Supervisor/Operator Performing the Verification Procedure:
 Name: Trp. J. Santora ID# 2276 ✓ Date: 1/11/2011 ✓

A Agency: Alaska State Troopers Phone # 656-1233

Instrument Location: Galena Police Department

B Alco S/N: 78855 ✓ Target Value: .081 ✓ High Pressure: 1000

Alco Test Values .084 ✓ .084 ✓
 1st Alco 2nd Alco

Signature [Handwritten Signature]
 (OVER)

[Handwritten initials]
 2/8/11

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Handwritten Signature]
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program



Subscribed and sworn before me this 11th day of MAR, 2011 *[Handwritten]*

[Handwritten Signature] (Notary Seal Stamp)
 Daniel L. Nicholas
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130135

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130135

JANUARY 11, 2011

OPERATOR'S NAME: SANTORA JEANNINE L
OPERATOR'S NUMBER: 2276
SUBJECT'S LAST NAME: VERIFICATION
SUBJECT'S FIRST NAME/NI :
OF CAL

O.L. #: 0000000
DEPT/AGENCY: GAL0
CASE/REPORT:
TEST TYPE: U
ALCO TARGET VALUE: .001
ALCO S/N: 78855

--- BREATH ANALYSIS ---

.001 ADJUSTED FOR 31.05 in		
ALCO TARGET	.004	15:29
BLANK TEST	.000	15:30
INTERNAL STANDARD	VERIFIED	15:31
ALCO TU 31.05 in	.004	15:31
BLANK TEST	.000	15:32
SUBJECT SAMPLE	.000	15:32
BLANK TEST	.000	15:33
ALCO TU 31.05 in	.004	15:33
BLANK TEST	.000	15:34

BMB 2/8/11

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130135

JANUARY 11, 2011
TIME 15:35

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	40c
BREATH TUBE:	43c
BAROMETER:	31.06 in
FLOW DETECTOR:	OKAY
PUMP HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!@#%&'()*+,-./0123456789:;<=>?@BCDEF6
HIJKL/MNOPQRSTUVWXYZ*~! " abcdefghijklmnop
qrstuvwxyz{|}~