

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DEC 03 2011

DataMaster cdm S/N: 130135

Supervisor/Operator Performing the Verification Procedure:

Name: Trp. J. Santora ID# 2276 Date: 12/1/2011

A

Agency: Alaska State Troopers Phone # 656-1233

Instrument Location: Galena Police Department

B

Alco S/N: 78855 Target Value: .081 High Pressure: 700

Alco Test Values .080 .080

1<sup>st</sup> Alco

2<sup>nd</sup> Alco

Signature

J. Santora  
(OVER)

CRP  
12/6/11

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 2<sup>nd</sup> day of Feb, 2012 NJK 2010.

Nikki Roth  
(Notary Seal Stamp)  
Daniel L. Nicholas Nikki Roth NJK  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130135

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130135

DECEMBER 01, 2011 *COB 12/6/11*

OPERATOR'S NAME:  
SANTORA JEANNINE L  
OPERATOR'S NUMBER: 2276  
SUBJECT'S LAST NAME:  
VERIFICATION  
SUBJECT'S FIRST NAME/NI :  
OF CAL  
O.L. #: 0000000  
DEPT/AGENCY: GALI  
CASE/REPORT:  
TEST TYPE: U  
ALCO TARGET VALUE: .001  
ALCO S/N: 78855

--- BREATH ANALYSIS ---

.001 ADJUSTED FOR 29.83 in		
ALCO TARGET	.000	15:00
BLANK TEST	.000	15:01
INTERNAL STANDARD	VERIFIED	15:01
ALCO TO 29.83 in	.000	15:01
BLANK TEST	.000	15:02
SUBJECT SAMPLE <i>COB 12/6/11</i>	.000	15:03
BLANK TEST	.000	15:04
ALCO TO 29.83 in	.000	15:05
BLANK TEST	.000	15:05

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130135

DECEMBER 01, 2011  
TIME 15:06

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	50c
BREATH TUBE:	41c
BAROMETER:	29.83 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST  
!@#%&'()\*+,-./0123456789:;<=>?@ABCDEF6  
HIJKL NOPQRSTU VWXYZ[\]^\_`abcde fghijklmnop  
qrstuvwxy z{|}~