

DEC 20 2010

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130136 ✓
Name	Daniel G. Gunderson	ID#	6602 ✓
		Date	12-14-10 ✓
A	Agency	AST	Phone # 543-2294
	Instrument Location	BETHEL AST	
B	Alco S/N	X124623 ✓	Target Value .081 ✓
			High Pressure 1150
	Alco Test Values	.085 ✓	.085 ✓
		1 st Alco	2 nd Alco
	Signature	[Signature]	
			COB 12/30/10

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 15th day of JAN 2011

[Signature]
 (Notary Seal Stamp)
 Carolyn M. Noland
 Notary Public, State of Alaska
 Commission Expires with Office

BT9 06/03



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130136

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130136

DECEMBER 14, 2010

OPERATOR'S NAME: *COB 12/20/10*
SUNDERSON/DANIEL E
OPERATOR'S NUMBER: 4502
SUBJECT'S LAST NAME:
TEST
SUBJECT'S FIRST NAME/INITIAL:
TEST
D.L. #: 00
DEPT/AGENCY: BETA
CASE/REPORT: 00
TEST TYPE: U
ALCO TARGET VALUE: .061
ALCO S.M. 10124517

--- BREATH ANALYSIS ---

.201 ADJUSTED FOR 30.86 in
ALCO TARGET .061 10141
BLANK TEST .000 10142
INTERNAL STANDARD VERIFIED 10142
ALCO TV 30.86 in .065 10143
BLANK TEST .000 10144
SUBJECT SAMPLE *COB 12/30/10* .000 10145
BLANK TEST .000 10146
ALCO TV 30.87 in .065 10147
BLANK TEST .000 10148

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130136

DECEMBER 14, 2010
TIME 10:18

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/09/07
HEATERS
SAMPLE CHAMBER: 50c
BREATH TUBE: 42c
BAROMETER: 30.87 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

ALCO TV 30.86 in .065 10143
ALCO TV 30.87 in .065 10147