

JUN 13 2011

VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm Breath Test Instrument
State of Alaska*

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130141</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>Scott Gardner</u>	ID# <u>6222</u> Date <u>16 May 2011</u>
A Agency <u>JBER4 Joint Base Elmendorf-Richardson</u>	Phone # <u>907-552-3537</u>
Instrument Location <u>Bldg 8517 20th St Joint Base Elmendorf-Richardson</u>	
B Alco S/N <u>A090740</u> Target Value <u>.083</u> High Pressure <u>500</u>	
Alco Test Values	<u>.081</u> <u>.085</u> 1 st Alco 2 nd Alco
Signature <u><i>Scott Gardner</i></u>	

(OVER)

*BMB
6/15/11*

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 16 day of June, 2011. NJB

Kyle Klopfer (Notary Seal Stamp)
 Daniel L. Nicholas KYLE KLOPPER
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130141

BMB
6/15/11

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130141

MAY 16, 2011

OPERATOR'S NAME:
GARONER/SCOTT/E
OPERATOR'S NUMBER: 6222
SUBJECT'S LAST NAME:
N/A

SUBJECT'S FIRST NAME/MI :
N/A

D.L. #: VERIFICATION OF CALIBRATION
DEPT/AGENCY: JBR4
CASE/REPORT: N/A
TEST TYPE: U
ALCO TARGET VALUE: .083
ALCO S/N: A090740

— BREATH ANALYSIS —

.083 ADJUSTED FOR 29.62 in
ALCO TARGET .082 08:31
BLANK TEST .000 08:32
INTERNAL STANDARD VERIFIED 08:32
ALCO TV 29.62 in .081 08:33
BLANK TEST .000 08:33
SUBJECT SAMPLE .000 08:34
BLANK TEST .000 08:35
ALCO TV 29.62 in .085 08:35
BLANK TEST .000 08:36

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130141

MAY 16, 2011
TIME 08:22

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 41c
BAROMETER: 29.62 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxy z{|}~■

