

MAY 17 2011

VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm Breath Test Instrument
State of Alaska*

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130142</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>Lynne Gralewski-Baugus</u>	ID# <u>6787</u> Date <u>5/3/11</u> <i>BMB 5/18/11</i>
A Agency <u>Joint Base Elmendorf-Richardson</u> Phone # <u>907-552-4000</u>	
Instrument Location <u>Bldg 656 A St JBER, AK 99506</u>	
B Alco S/N <u>X301548</u> Target Value <u>0.076</u> High Pressure <u>1100 psi</u>	
Alco Test Values	<u>.074</u> <u>.073</u> 1st Alco 2 nd Alco
Signature	<u><i>Lynne Gralewski-Baugus</i></u>

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 15 day of JUNE, 2011 *2011 NJB*

Kyle Klopfer (Notary Seal Stamp)
 Daniel L. Nicholas *KYLE KLOPPER*
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130142

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130142

MAY 03, 2011

OPERATOR'S NAME: *BMB 5/18/11*
GRALSKI-BAUSUS/LYNE/MARIE
OPERATOR'S NUMBER: 8787
SUBJECT'S LAST NAME:
VERIFICATION OF CALIBRATION
SUBJECT'S FIRST NAME/NI :
VOC
O.L. #: VOC
DEPT/AGENCY: EDF4
BASE/REPORT: VOC
TEST TYPE: V
ALCO TARGET VALUE: 0.076
ALCO S/N: X301548

--- BREATH ANALYSIS ---

.076 ADJUSTED FOR 29.75 in
ALCO TARGET .075 09:10
BLANK TEST .000 09:11
INTERNAL STANDARD VERIFIED 09:11
ALCO TO 29.75 in .074 09:13
BLANK TEST .000 09:14
SUBJECT SAMPLE *BMB 5/18/11* .000 09:15
BLANK TEST .000 09:16
ALCO TO 29.75 in .073 09:17
BLANK TEST .000 09:17

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130142

MAY 03, 2011
TIME 09:02

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 07-09-06
HEATERS
SAMPLE CHAMBER: 45c
BREATH TUBE: 45c
BAROMETER: 29.75 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#\$%^&*()~ - #0123456789:;<=>?@BCDEFG
HIJKL^NOPQRSTUWVXYZ\|_`abcedefghijklmno
pqrstuvwxyz{|}-||

