

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130142</u> ✓
Name <u>Dickie R. Powell</u>		ID# <u>1168</u> Date <u>12 Feb 2009</u> ✓
A	Agency <u>Directorate of Emergency Service</u>	Phone # <u>384-0812</u>
Instrument Location <u>Bldg# 656, A Street, Fort Richardson, Alaska 99505</u> ✓		
Alco S/N <u>X172969</u> ✓		Target Value <u>.079</u> ✓ High Pressure <u>400 psi</u>
B	Alco Test Values	
	<u>.087</u> 1 st Alco	<u>.086</u> 2 nd Alco
Signature <u></u>		<u>CS</u> <u>2/23/09</u>
(OVER)		


Under the Alaska Rules of Evidence, I certify that;

(Do Not write in the area below)


(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Dickie R Powell, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,



Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

02/23/09

Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130142 ✓

(CONTINUED FROM FRONT PAGE)

05/23/09

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130142

FEBRUARY 12, 2009

OPERATOR'S NAME: POWELL, DICKIE R
OPERATOR'S NUMBER: 1168
SUBJECT'S LAST NAME: VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI : N/A

O.L. #: N/A
DEPT/AGENCY: FRM
CASE/REPORT: N/A
TEST TYPE: U
ALCO TARGET VALUE: .079 ✓
ALCO S/N: X172969 ✓

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 30.19 in
ALCO TARGET .079 16:14
BLANK TEST .000 16:15
INTERNAL STANDARD VERIFIED 16:16
ALCO TV 30.19 in .087 ✓ 16:16
BLANK TEST .000 16:17
SUBJECT SAMPLE .000 16:17
BLANK TEST .000 16:18
ALCO TV 30.19 in .086 ✓ 16:18
BLANK TEST .000 16:19

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130142

FEBRUARY 12, 2009 ✓
TIME 16:19

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 03/29/00
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 44c
BAROMETER: 30.19 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ ¡ ¢