

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAR 11 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130142</u>
Name <u>Dickie R. Powell</u>		ID# <u>1168</u> Date <u>09 Mar 2009</u>
Agency <u>Directorate of Emergency Service</u>		Phone # <u>384-0812</u>
Instrument Location <u>Bldg# 656, A Street, Fort Richardson, Alaska 99505</u>		
Alco S/N <u>78820</u>		Target Value <u>.079</u> High Pressure <u>1100 psi</u>
Alco Test Values		
	<u>.082</u> 1 <sup>st</sup> Alco	<u>.082</u> 2 <sup>nd</sup> Alco
Signature <u><i>Dickie R Powell</i></u>		<u>CSB</u> <u>3/24/09</u>
(OVER)		

Under the Alaska Rules of Evidence, I certify that;  
(Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Dickie R Powell, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



*Jeanne Swartz* 3/24/09  
Date  
Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130142 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130142

MARCH 09, 2009

OPERATOR'S NAME: POWELL DICKIE R  
OPERATOR'S NUMBER: 1168  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/MI :  
N/A

O.L. #: N/A  
DEPT/AGENCY: FRM  
CASE/REPORT: N/A  
TEST TYPE: U  
ALCO TARGET VALUE: .079 ✓  
ALCO S/N: 78820 ✓

### --- BREATH ANALYSIS ---

.079 ADJUSTED FOR 30.04 in  
ALCO TARGET .079 14:41  
BLANK TEST .000 14:42  
INTERNAL STANDARD VERIFIED 14:42  
ALCO TV 30.04 in .082 ✓ 14:42  
BLANK TEST .000 14:43  
SUBJECT SAMPLE .000 14:44  
BLANK TEST .000 14:45  
ALCO TV 30.04 in .082 ✓ 14:45  
BLANK TEST .000 14:46

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130142

MARCH 09, 2009  
TIME 14:46

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 03/29/00  
HEATERS  
SAMPLE CHAMBER: 48c  
BREATH TUBE: 43c  
BAROMETER: 30.04 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### · PRINTER TEST ·

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEF  
HIJKL MNOPQRSTUVWXYZ[\]^\_`abcde fghijklmno  
pqrstuvwxy z{|}~#