

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130142</u> ✓
Name	<u>Dick Powell</u>	ID#	<u>1168</u> ✓
Date	<u>19 NOV 08</u> ✓		
Agency	<u>FRN4 FT Richardson, AK</u>	Phone #	<u>384-0812</u>
Instrument Location	<u>Bldg # 656 A ST, FT Richardson, AK</u>		
Alco S/N	<u>X172969</u> ✓	Target Value	<u>.079</u>
High Pressure	<u>850 psi</u>		
Alco Test Values	<u>.085</u> ✓	<u>.086</u> ✓	
1 st Alco		2 nd Alco	
Signature	<u>Dick R Powell</u> ✓		
(OVER)			

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Dickie R Powell, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

Jeanne Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

11/25/08

Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130142 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130142

NOVEMBER 19, 2008 ✓ *J 11/25/08*

OPERATOR'S NAME:
POWELL, DICKIE R
OPERATOR'S NUMBER: 1168
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
N/A
O.L. #: N/A
DEPT/AGENCY: FRM4
CASE/REPORT: N/A
TEST TYPE: U
ALCO TARGET VALUE: .079 ✓
ALCO S/N: X172969

— BREATH ANALYSIS —

.079 ADJUSTED FOR 29.94 in		
ALCO TARGET	.079	15:31
BLANK TEST	.000	15:32
INTERNAL STANDARD	VERIFIED	15:32
ALCO TV 29.94 in	.085 ✓	15:33
BLANK TEST	.000	15:33
SUBJECT SAMPLE	.000	15:34
BLANK TEST	.000	15:35
ALCO TV 29.94 in	.086 ✓	15:35
BLANK TEST	.000	15:36

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130142

NOVEMBER 19, 2008 ✓
TIME 15:37

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	03/29/00
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	44c
BAROMETER:	29.94 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!@#%&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmno
pqrstuvwxyz{|}~■