

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DEC 03 2009

DataMaster cdm S/N <u>130142</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>Scott E. Gardner</u>	ID# <u>6222</u> Date <u>30 Nov 09</u>
A Agency <u>Directorate of Emergency Services</u> Phone # <u>384-3888</u>	
Instrument Location <u>Bldg #656, A St, Ft Richardson, AK 99505</u>	
B Alco S/N <u>124656</u> Target Value <u>.080</u> High Pressure <u>650 psi</u>	
Alco Test Values <u>.081</u> 1 st Alco <u>.081</u> 2 nd Alco	
Signature <u><i>Scott E. Gardner</i></u>	
(OVER)	


(Do Not write in the area below)

- I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
 - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
 - (3) I am the Scientific Director of the State Breath Alcohol Program.
 - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
 - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
 - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 31st day of December, 2009.

Carolyn M. Noland (Notary Seal Stamp)
 Carolyn M. Noland
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130142 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130142

NOVEMBER 30, 2009 ✓ *CS 12/1/09*

OPERATOR'S NAME:
GARDNER/SCOTT/E
OPERATOR'S NUMBER: 6222
SUBJECT'S LAST NAME:
VERIFY CAL
SUBJECT'S FIRST NAME/MI :
N/A

O.L. #: N/A
DEPT/AGENCY: FRM
CASE/REPORT: N/A
TEST TYPE: U
ALCO TARGET VALUE: .080 ✓
ALCO S/N: 124656 ✓

NTP
12/30/09 X BREATH ANALYSIS ---

.080 ADJUSTED FOR 29.42 in ✓
ALCO TARGET .078 14:14
BLANK TEST .000 14:15
INTERNAL STANDARD VERIFIED 14:15
ALCO TU 29.42 in .081 14:15 ✓
BLANK TEST .000 14:16
SUBJECT SAMPLE *CS 12/1/09* .000 14:17
BLANK TEST .000 14:18 ✓
ALCO TU 29.42 in .081 14:18
BLANK TEST .000 14:19

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130142

NOVEMBER 30, 2009 ✓
TIME 14:11

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 03/29/00
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 46c
BAROMETER: 29.42 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKL NOPQRSTU VWXYZ[\]^_`abcde fghijklmno
pqrstuvwxy z{|}~■