

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

JUL 13 2010

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130142 ✓
Name	Scott E. Gardner	ID#	6222 ✓
Date	6 Jul 10		✓
A	Agency	Directorate of Emergency Services	Phone # 384-3888
	Instrument Location	Bldg #656, A St, Ft Richardson, AK 99505	
B	Alco S/N	X301542 ✓	Target Value .082 ✓
		High Pressure	300 psi
	Alco Test Values	.085 ✓	.086 ✓
		1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
	Signature		
			COB 7/15/10

(OVER)


(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 27<sup>th</sup> day of July, 2010.

  
(Notary Seal Stamp)

Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130142

(CONTINUED FROM FRONT PAGE)

**C** TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130142

JULY 06, 2010

OPERATOR'S NAME:

GARDNER/SCOTT/E

OPERATOR'S NUMBER: 6222

SUBJECT'S LAST NAME:

N/A

SUBJECT'S FIRST NAME/MI :

N/A

O.L. #: VERIFICATION OF CAL

DEPT/AGENCY: FRM4

CASE/REPORT: N/A

TEST TYPE: U

ALCO TARGET VALUE: .082

ALCO S/N: K301542

COB 7/15/10

--- BREATH ANALYSIS ---

.082 ADJUSTED FOR 30.06 in  
ALCO TARGET .082 14:54  
BLANK TEST .000 14:55  
INTERNAL STANDARD VERIFIED 14:55  
ALCO TV 30.06 in .085 14:56  
BLANK TEST .000 14:57  
SUBJECT SAMPLE COB 7/15/10 .000 14:57  
BLANK TEST .000 14:58  
ALCO TV 30.06 in .086 14:58  
BLANK TEST .000 14:59

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130142

JULY 06, 2010

TIME 14:51

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 03/29/00

HEATERS

SAMPLE CHAMBER: 48c

BREATH TUBE: 44c

BAROMETER: 30.06 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~ ¡ ¢