

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

FEB 23 2009

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130144</u>
Name	<u>J Swartz</u>	ID#	<u>4314</u> Date <u>2/21/09</u>
A Agency	<u>SCDL</u>	Phone #	<u>269-5592</u>
Instrument Location	<u>SCDL</u>		
B Alco S/N	<u>SC0708</u>	Target Value	<u>.080</u> High Pressure <u>900 psi</u>
Alco Test Values	<u>.082</u> 1 <sup>st</sup> Alco	<u>.082</u> 2 <sup>nd</sup> Alco	
Signature	<u>J Swartz</u>		<u>CAB</u> <u>3/5/09</u>

(OVER)

Under the Alaska Rules of Evidence, I certify that (write in the area below)

- (a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.
- (b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Jeanne M Swartz, whose name appears in my records, is qualified to verify the calibration of the above instrument.
- (c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



J Swartz

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

3/2/09  
Date

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130144 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130144 ✓

FEBRUARY 21, 2009 ✓

OPERATOR'S NAME:  
SMARTZ/J

OPERATOR'S NUMBER: 4314

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/NI :

L

O.L. #: L

DEPT/AGENCY: AAAL

CASE/REPORT:

TEST TYPE: U

ALCO TARGET VALUE: .000 ✓

ALCO S/N: SCOTT08 ✓

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 30.63 in  
ALCO TARGET .001 07:06  
BLANK TEST .000 07:07  
INTERNAL STANDARD VERIFIED 07:07  
ALCO TV 30.63 in .002 ✓ 07:08  
BLANK TEST .000 07:09  
SUBJECT SAMPLE .000 07:10  
BLANK TEST .000 07:10  
ALCO TV 30.63 in .002 ✓ 07:11  
BLANK TEST .000 07:12

2/21/09

J 2/21/09

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130144 ✓

FEBRUARY 21, 2009 ✓  
TIME 07:04

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS:

SAMPLE CHAMBER: 50c

BREATH TUBE: 41c

BAROMETER: 30.63 in

FLOW DETECTOR: OKAY

PUMP:

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@ABCDEF  
HIJKLNMOPQRSTUVWXYZ[\]^\_`ab|cdefghi jk lmnop  
qrs|tuvwxyz{|}~■