

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

AUG 25 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

| | | |
|--|--------------------------------|--|
| Supervisor/Operator Performing the Verification Procedure: | | DataMaster cdm S/N <u>130144</u> ✓ |
| Name <u>RAY CONTRERAS</u> | | ID# <u>6350</u> Date <u>3/16/09</u> ✓ |
| A | Agency <u>3 SFS</u> | Phone # <u>552-6567</u> |
| Instrument Location <u>ELMENDORF AFB</u> ✓ | | |
| Alco S/N <u>X172959</u> ✓ | | Target Value <u>.081</u> High Pressure _____ |
| B | Alco Test Values <u>.081</u> ✓ | <u>.081</u> ✓ |
| | 1 st Alco | 2 nd Alco |
| Signature <u>Ray Contreras</u> ✓ | | <u>CRB</u> <u>8/21/09</u> |

(OVER)

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 4th day of Sept., 2009.

Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130144

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

Breath Test Record

Serial Number Model
Location

03/16/09

Diagnostic Test Record

Serial Number Model
Location

Subject Information

Last Name O.L. #
First Name

Operator Information

Operator Name Case
Operator # Department

Test Results

Alco Target Alco Serial
Nominal Barometer Test Type
03/16/09 Dry Gas 1 Time: Test Date
Subject Alcohol Time: Test Time
Dry Gas 2 Time: Report

Diagnostic Test Results

Test Date ✓ Computer
Test Time Program
Software Date
Sample Chamber (C)
Breath Tube (C)
Barometer (in)
Flow Detector
High Speed Pump
Detector
Filter
Quartz
Calibration