


# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130144</u> ✓
Name	<u>Jose Ramirez</u>	ID# <u>6532</u> Date <u>24 Apr 08</u> ✓
A	Agency <u>3rd Security Forces Squadron</u>	Phone # <u>552-6570</u>
Instrument Location <u>Elmendorf AFB, AK, Law Enforcement Desk</u>		
Alco S/N <u>X301568</u> ✓		Target Value <u>.074</u> ✓ High Pressure <u>1250</u>
Alco Test Values		
	<u>.082</u> ✓ 1 <sup>st</sup> Alco	<u>.081</u> ✓ 2 <sup>nd</sup> Alco
Signature	<u></u>	<u>CAS</u> <u>5/5/08</u>
(OVER)		

Under the Alaska Rules of Evidence, I certify that, (Print name in the area below)


(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Jose A Ramirez, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,





Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

5/5/08  
Date

APR 25 2008

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130144-8 5/5/08  
~~X301568~~

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130144

APRIL 24, 2008

OPERATOR'S NAME: *COB 5/5/08*  
RAMIREZ/JOSE/A  
OPERATOR'S NUMBER: 6532  
SUBJECT'S LAST NAME:  
VOC  
SUBJECT'S FIRST NAME/MI:  
VOC  
O.L. #: 1234567890  
DEPT/AGENCY: EDF4  
CASE/REPORT: 080423  
TEST TYPE: V  
ALCO TARGET VALUE: .074  
ALCO S/N: X301568

--- BREATH ANALYSIS ---

.074 ADJUSTED FOR 29.99 in		
ALCO TARGET	.074	16:10
BLANK TEST	.000	16:11
INTERNAL STANDARD	VERIFIED	16:12
ALCO TV 29.99 in	.082	16:12
BLANK TEST	.000	16:13
SUBJECT SAMPLE	.000	16:13
BLANK TEST	.000	16:14
ALCO TV 29.99 in	.081	16:14
BLANK TEST	.000	16:15

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130144

APRIL 24, 2008  
TIME 15:23

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS /	
SAMPLE CHAMBER:	50c
BREATH TUBE:	42c
BAROMETER:	29.99 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@ABCDEF  
HIJKL@PQRSTUWXYZ\]^\_`abcde fghi jk lmn  
opqrstuvwxyz{|}~#