

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

APR 30 2010

Supervisor/Operator Performing the Verification Procedure:

DataMaster cdm S/N 130144

Name CHRISTOPHER M. CHAPMAN

ID# 6390

Date

27 Apr 10

18 APR 10 ⁰⁰⁸

5/1/10

A

Agency EDF4 ELMENDORF AFB

Phone # 552-0176

Instrument Location 8517 20TH STREET ELMENDORF AFB, AK

B

Alco S/N X124658

Target Value .081

High Pressure 900

Alco Test Value Average .071

1st Alco

.072

2nd Alco

Signature

(OVER)

Handwritten initials and date:
bmb
5/3/10

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Handwritten signature of Nita J. Bolz

Nita J. Bolz

Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 07th day of June, 2010.

Handwritten signature of Carolyn M. Noland

(Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130144

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130144

APRIL 27, 2010

OPERATOR'S NAME:

CHAPMAN CHRISTOPHER M

OPERATOR'S NUMBER: 6390

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 1

DEPT/AGENCY: EDF4

CASE/REPORT:

TEST TYPE: U

ALCO TARGET VALUE: .081

ALCO S/N: X124658

— BREATH ANALYSIS —

.081 ADJUSTED FOR 29.36 in

ALCO TARGET	.079	14:54
BLANK TEST	.000	14:55
INTERNAL STANDARD	VERIFIED	14:55
ALCO TV 29.36 in	.071	14:55
BLANK TEST	.000	14:56
SUBJECT SAMPLE	.000	14:57
BLANK TEST	.000	14:57
ALCO TV 29.36 in	.072	14:58
BLANK TEST	.000	14:58

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130144

APRIL 27, 2010

TIME 15:00

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 42c

BAROMETER: 29.36 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ ¡ ¢