

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130144

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130144

SEPTEMBER 29, 2009

OPERATOR'S NAME:
CHAPMAN CHRISTOPHER M
OPERATOR'S NUMBER: 6390
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI:
VOC

O.L. #: 1
DEPT/AGENCY: EDF4
CASE/REPORT:
TEST TYPE: V

ALCO TARGET VALUE: .081
ALCO S/N: 124669

— BREATH ANALYSIS —

.081 ADJUSTED FOR 29.78 in
ALCO TARGET .080 07:22
BLANK TEST .080 07:23
INTERNAL STANDARD VERIFIED 07:23
ALCO TV 29.78 in .082 07:23
BLANK TEST .080 07:24
SUBJECT SAMPLE .080 07:25
BLANK TEST .080 07:26
ALCO TV 29.78 in .082 07:26
BLANK TEST .080 07:27

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130144

SEPTEMBER 29, 2009
TIME 07:28

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.78 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#\$%^&*+,-./0123456789:;<=>?@BCDEFG
HIJKLNPQRSTUUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~!@

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

OCT 26 2009


Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N 130144
Name: Christopher Chapman	ID# 6390	Date 29 Sep 2009
A Agency 3 rd Security Forces Squadron		Phone # (907)552-0176
Instrument Location: Law Enforcement desk, Elmendorf AFB, AK		
B Alco S/N X172959		Target Value .081
		High Pressure 600psi
Alco Test Values	0.082	0.082
	1 st Alco	2 nd Alco
Signature		29 Sep 09
(OVER)		

BMB
11/14/09


(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 3rd day of Dec, 2009.


(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

