

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm Breath Test Instrument  
State of Alaska*

**Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program**

NOV 20 2009

<b>DataMaster cdm S/N 130144</b>	
Supervisor/Operator Performing the Verification Procedure:	
Name: Jose Ramirez	ID# 6532
Date 27 Oct 2009 ✓	
Agency 3 <sup>rd</sup> Security Forces Squadron	Phone # (907)552-9863
Instrument Location: Law Enforcement desk, Elmendorf AFB, AK ✓	
Alco S/N X301542 ✓ Target Value .080 High Pressure 1290psi	
Alco Test Values	0.082 1 <sup>st</sup> Alco
	0.082 2 <sup>nd</sup> Alco
Signature _____	COB 12/7/09
(OVER)	

(Do not write in the area below)

- I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
  - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
  - (3) I am the Scientific Director of the State Breath Alcohol Program.
  - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
  - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
  - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

*Nita J. Bolz*  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 21<sup>st</sup> day of Dec., 2009.

*Carolyn M. Noland*  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130144 ✓

(CONTINUED FROM FRONT PAGE)

**C** TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130144 ✓

OCTOBER 27, 2009

OPERATOR'S NAME:  
POMPEZ/JOE/A  
OPERATOR'S NUMBER: 6502  
SUBJECT'S LAST NAME:  
UOC

SUBJECT'S FIRST NAME/MI :  
UOC

D.L. #: 1274567890  
DEPT/AGENCY: ED-4  
CASE/REPORT: 20091027  
TEST TYPE: T ✓  
ALCO TARGET VALUE: .080 ✓  
ALCO S/N: A361512 ✓

### --- BREATH ANALYSIS ---

.080 ADJUSTED FOR 30.04 in  
ALCO TARGET .080 12:56  
BLANK TEST .000 12:57  
INTERNAL STANDARD VERIFIED 12:57  
ALCO TU 30.06 in .082 ✓ 12:57  
BLANK TEST .000 12:58  
SUBJECT SAMPLE .000 12:59  
BLANK TEST .000 13:00  
ALCO TU 30.06 in .082 ✓ 13:00  
BLANK TEST .000 13:01

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130144 ✓

OCTOBER 27, 2009 ✓  
TIME 13:02

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 01/08/09  
HEATERS  
SAMPLE CHANGER: 49c  
BREATH TUBE: 49c  
BAROMETER: 30.06 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY  
PRINTER TEST  
ALCO IN PROGRESS: 0.000  
ALCO S/N: A361512