

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DEC 21 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130171</u>
A	Name <u>Gregory S. Larmore</u>	ID# <u>691</u> Date <u>3/20/07</u>
	Agency <u>E.14</u>	Phone # <u>283</u>
	Instrument Location <u>Rm 143</u>	
B	Alco S/N <u>58428</u>	Target Value <u>.080</u> High Pressure <u>29.80</u>
	Alco Test Values <u>0.85</u> ✓ <u>0.85</u> ✓	1 <sup>st</sup> Alco 2 <sup>nd</sup> Alco
	Signature <u>[Signature]</u>	

(OVER)

BMB  
1/5/10

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 20 day of January, 2010.

[Signature] (Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130171 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

<b>Breath Test Record</b>		
Serial Number <u>130171</u>	Model <u>DataMaster</u>	
Location <u>EIL4</u>	<u>Eielson AFB/354 SFS</u>	
<b>Subject Information</b>		
Last Name <u>Voc</u>	O.L. # <u>V</u>	
First Name <u>Voc</u>		
<b>Operator Information</b>		
Operator Name <u>Lormore</u>	Case <u>V</u>	
Operator # <u>6691</u>	Department <u>EIL4</u>	
<b>Test Results</b>		
Alco Target <u>.080</u>	Alco Serial <u>58428</u>	
Nominal <u>0.079</u>	Barometer <u>29.80</u>	Test Type <u>V</u>
Dry Gas 1 <u>0.085</u>	Time: <u>15:46:15</u>	Test Date <u>03/20/2009</u>
Subject Alcohol <u>0.000</u>	Time: <u>15:47:32</u>	Test Time <u>15:44:35</u>
Dry Gas 2 <u>0.085</u>	Time: <u>15:48:53</u>	Report <u>Y</u>

## DIAGNOSTIC CHECK

<b>Diagnostic Test Record</b>	
Serial Number <u>130171</u>	Model <u>DataMaster</u>
Location <u>EIL4</u>	<u>Eielson AFB/354 SFS</u>
<b>Diagnostic Test Results</b>	
Test Date <u>03/20/2009</u> ✓	Computer <u>Okay</u>
Test Time <u>15:31:37</u>	Program <u>Okay</u>
<u>BMB</u> <u>1/5/10</u>	Software Date <u>02/20/2001</u>
	Sample Chamber (C) <u>49</u>
	Breath Tube (C) <u>42</u>
	Barometer (in) <u>29.80</u>
	Flow Detector <u>Okay</u>
	High Speed Pump <u>Okay</u>
	Detector <u>Okay</u>
	Filter <u>Okay</u>
	Quartz <u>Okay</u>
	Calibration <u>Okay</u>