

VERIFICATION OF CALIBRATION REPORT
of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program


DataMaster cdm S/N <u>130171</u>		
Supervisor/Operator Performing the Verification Procedure:		
Name <u>CHRISTOPHER CHAPMAN</u>	ID# <u>6390</u> Date <u>20 Jun 08</u> ✓	
A Agency <u>EILY</u>	Phone # <u>377-1408</u>	
Instrument Location <u>451 DIVISION ST. EIELSON AFB, AK 99702</u> ✓		
B Alco S/N <u>X301542</u> ✓ Target Value <u>.075</u> ✓ High Pressure <u>890</u>		
Alco Test Values	<u>.079</u> ✓ 1 st Alco	<u>.079</u> ✓ 2 nd Alco
Signature <u></u>	<u>COB</u> <u>7/7/08</u>	
(OVER)		

Under the Alaska Rules of Evidence, I certify that;

(Do Not write in the area below)

- (a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.
- (b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Christopher M Chapman, whose name appears in my records, is qualified to verify the calibration of the above instrument.
- (c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



 07/09/08
 Jeanne Swartz Date
 Scientific Director Designee
 State of Alaska
 State Breath Alcohol Testing Program

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130171 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130171

JUNE 20, 2008 ✓ *COS 7/7/08*

OPERATOR'S NAME:
CHAPMAN CHRISTOPHER M
OPERATOR'S NUMBER: 6390
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
UOC
O.L. #: 1
DEPT/AGENCY: EIL4
CASE/REPORT:
TEST TYPE: U
ALCO TARGET VALUE: .075 ✓
ALCO S/N: X301542 ✓

— BREATH ANALYSIS —

.075 ADJUSTED FOR 29.57 in
ALCO TARGET .074 15:50
BLANK TEST .000 15:51
INTERNAL STANDARD VERIFIED 15:51
ALCO TV 29.57 in .079 ✓ 15:52
BLANK TEST .000 *COS 7/7/08* 15:53
SUBJECT SAMPLE .000 15:53
BLANK TEST .000 15:54
ALCO TV 29.57 in .079 15:54
BLANK TEST .000 15:55

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130171

JUNE 20, 2008 ✓
TIME 15:58

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.57 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; (=) ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~