

VERIFICATION OF CALIBRATION REPORT
of DataMaster cdm Breath Test Instrument
 State of Alaska
 Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program


Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130171</u>
Name <u>Chris Howard</u>	ID# <u>6581</u>	Date <u>5-NOV-2008</u>
A Agency <u>ETC</u>		Phone # <u>9073775224</u>
Instrument Location <u>457 DIVISION ST, EISENHARDT AFB AK 99702</u>		
B Alco S/N <u>X301542</u>		Target Value <u>.075</u> High Pressure <u>500-510</u>
Alco Test Values	<u>1.080</u> 1 st Alco	<u>.079</u> 2 nd Alco
Signature <u>[Signature]</u>		<u>CS</u> <u>11/18/08</u>

(OVER)

Under the Alaska Rules of Evidence, I certify that:

(Do Not write in the area below)

- (a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.
- (b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Chris M Howard, whose name appears in my records, is qualified to verify the calibration of the above instrument.
- (c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

[Signature]

Jeanne Swartz
 Scientific Director Designee
 State of Alaska
 State Breath Alcohol Testing Program

11/19/08
 Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130171 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130171

NOVEMBER 05, 2008 *CPB 11/05/08*

OPERATOR'S NAME:
HOWARD/CHRIS/H
OPERATOR'S NUMBER: 6581
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
L
O.L. #: L
DEPT/AGENCY: EIL4
CASE REPORT:
TEST TYPE: V
ALCO TARGET VALUE: .075 ✓
ALCO S/N: X301542

--- BREATH ANALYSIS ---

.075 ADJUSTED FOR 29.59 in		
ALCO TARGET	.074	10:06
BLANK TEST	.000	10:06
INTERNAL STANDARD	VERIFIED	10:07
ALCO TV 29.59 in	.080 ✓	10:07
BLANK TEST	.000	10:08
SUBJECT SAMPLE	.000	10:08
BLANK TEST	.000	10:09
ALCO TV 29.59 in	.079 ✓	10:09
BLANK TEST	.000	10:10

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130171

NOVEMBER 05, 2008 ✓
TIME 10:01

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	29.59 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ ■